



705 Front Street
Toledo, OH 43605
Phone: 419-693-5307
Fax: 419-691-0418
www.lexamed.com

Skye Peptides

Lab # 2602311

PO # N/A

Test Article: Lyophilized peptide batch:
FOX26-15-001 Part # N/A Lot # FOX26 Batch #
15-001

USP <71> STERILITY TEST FINAL REPORT

Test article received: 3/20/2026
No. of articles tested: 20
Test start date: 3/20/2026
Test termination date: 4/6/2026
Procedure no: S-SRH-1-1
SOP No. (current version): LEXLP-017

Procedure: The outer package of each test article was disinfected prior to testing using a sterile clean room wipe moistened with a germicidal.

The samples were processed using the Membrane Filtration Method. Packages were opened aseptically. The contents of 20 samples were individually reconstituted then pooled and filtered through a 0.45 µm filter. The filter was cut in half, ½ was transferred directly to Fluid Thioglycollate Medium (FTM) and ½ to Soybean Casein Digest (SCD). A negative control was similarly prepared. The samples were incubated for a minimum of 14 days as indicated in the results table below.

The results are presented below and apply to the sample tested.

Results:

Test	Media	Media Volume (mL)	Incubation Temp (°C)	Number of Positives	Number of Negatives
½ Test filter	FTM	150	30-35	0	1
½ Test filter	SCD	100	20-25	0	1
½ Neg Control filter	FTM	150	30-35	0	1
½ Neg Control filter	SCD	100	20-25	0	1

Note: Per sponsor, the test was performed without a method suitability validation, therefore, results are for information only.

Record Storage: All raw data pertaining to this study will be maintained in the LexaMed archives for a minimum of 5 years.

Approved by

Jamara Allen

Tech: AL/DJF

Date 04/06/26



All reports are submitted as confidential communications. Reports may not be reproduced except in their entirety pending LexaMed approval.